



## Paper & Paperboard Testing Program

### Summary Report #3152 G - December 2021

---

[Introduction to the Paper & Paperboard Interlaboratory Program](#)

[Explanation of Tables and Definitions of Terms](#)

<b><u>Analysis</u></b>	<b><u>Analysis Name</u></b>
350	Color & Color Difference - Near White Papers - C/2deg obs
351	Color & Color Difference - Near White Papers - D65/10deg obs
360	Thickness (Caliper), Printing papers
361	Thickness (Caliper), Packaging papers
364	Coefficient of Static Friction - Horizontal Plane Method - Printing Papers
365	Coefficient of Kinetic Friction - Horizontal Plane Method - Printing Papers
370	Air Resistance - Gurley Oil Type
372	Porosity - Sheffield Type - Sheffield Units for 3/4 inch Diameter Orifice
376	Roughness - Print Surf Method - 0.5 to 4.0 Microns
377	Roughness - Print Surf Method - 2.5 to 6.0 Microns
378	Roughness - Sheffield Type
382	Moisture in Paper
384	Opacity (89% Reflectance Backing) - Fine Papers
386	Opacity (Paper Backing) - Fine Papers and Newsprint
390	Directional Brightness
391	Directional Brightness of Fluorescent Samples
392	Diffuse Brightness
394	Fluorescent Component of Directional Brightness
395	Specular Gloss at 75 Degrees - High Range
396	Specular Gloss at 75 Degrees - Low Range
398	Grammage (Mass per Unit Area)
399	Sizing Test (Hercules Type)

---

## **The CTS Paper & Paperboard Interlaboratory Program**

In 1969, the National Bureau of Standards (now designated the National Institute for Standards and Technology) and the Technical Association of the Pulp and Paper Industry (TAPPI) developed an interlaboratory program for paper and paperboard testing. Since 1971, Collaborative Testing Services has operated the Collaborative Reference Program for Paper and Paperboard. With hundreds of organizations from around the world participating in these tests, this program has become one of the largest of its kind. The program allows laboratories to compare the performance of their testing with that of other participating laboratories, and provides a realistic picture of the state of paper testing.

### **About CTS**

Founded in 1971, Collaborative Testing Services, Inc. (CTS) is a privately - owned company that specializes in interlaboratory tests for a variety of industrial sectors: rubber, plastics, fasteners and metals, CKPG, paper, color and wine, as well as proficiency tests for forensic laboratories. All of the tests are designed to assist organizations in achieving and maintaining quality assurance objectives. Labs from the U.S., as well as more than 80 countries, currently participate in CTS programs.

If there are any questions on the report or testing program, please contact:

**Collaborative Testing Services, Inc.**  
21331 Gentry Drive  
Sterling, Virginia 20166 USA  
+1-571-434-1925  
FAX #: +1-571-434-1937  
paper@cts-interlab.com

**Office Hours: 8:00 a.m. - 4:30 p.m. ET**

## Key for Web Summary Reports (Page 1 of 2)

<b>WebCode</b>	Assigned laboratory identification number (temporary) used to ensure lab confidentiality while permitting a lab to locate its data in the Paper Report published on the CTS Web site. The WebCode for each analysis can be found on the datasheets and in the Performance Analysis Report mailed to each participant.
<b>Lab Mean</b>	The average of the values obtained for each sample by the participant.
<b>Grand Mean</b>	The average of the LAB MEANS for all included participants. Laboratories flagged with an X or an M (see DATA FLAG column) are excluded from the GRAND MEAN.
<b>ΔE</b>	The calculated total color difference between the two samples. For the Hunter L,a,b analyses it is calculated in Hunter units (ΔE). For the L*,a*,b* analyses it is calculated in CIELAB units (ΔE*).
<b>Difference from Grand Mean</b>	The difference of the LAB MEAN from the GRAND MEAN.
<b>Between-Lab Standard Deviation</b>	An indication of the precision of measurement between the laboratories. The greater the spread of the LAB MEANS about the GRAND MEAN, the larger the BETWEEN-LAB STANDARD DEVIATION (and vice versa).
<b>Comparative Performance Value</b>	An indication of how well a laboratory's results agree with the other participants. The CPV is a ratio indicating the number of standard deviations from the GRAND MEAN. The closer a laboratory's COMPARATIVE PERFORMANCE VALUE is to zero, the more consistent its results are with the other participants' data (and vice versa). The critical value for each CPV will vary depending on the number of labs participating in a test.
<b>Inst Code</b>	A code indicating the manufacturer of the instrument used to perform the test (see separate INSTRUMENT CODE LIST for each test section), if instruments are tracked.
<b>Data Flag</b>	DATA FLAGS are assigned based on the simultaneous analysis of both samples tested. Refer to the following chart for an explanation of each symbol:

<u>DATA FLAG</u>	<u>STATISTICALLY INCLUDED/EXCLUDED</u>	<u>ACTION REQUIRED</u>
*	INCLUDED	<b>CAUTION</b> - review testing procedure and monitor future results. Results fall outside 95% ellipse but within a 99% ellipse that is calculated but not drawn.
X	EXCLUDED	<b>STOP</b> - immediate review of data and/or testing procedure is required. Results fall outside the 99% ellipse. See specific notes following each table for more information on why the data is excluded.
M	EXCLUDED	<b>PROCEED</b> - lab was unable to report data for at least one sample.

**Graph** - For each laboratory, the LAB MEAN for the first sample (x-axis) is plotted against the LAB MEAN for the second sample (y-axis) with each point representing a laboratory. The horizontal and vertical cross-hairs are the GRAND MEANS for each sample. When 20 or more laboratories are in the statistics, an ellipse is also drawn so that 95% of the time a randomly selected laboratory will be included inside the ellipse. Plotted data flags are explained on the previous page.

---

### Common Problems Highlighted in Footnotes

1. **Extreme data** - The laboratory's results for one or both samples are so inconsistent with those of the other participants that the lab mean(s) fall outside the plot. The participant is advised to immediately review his data and/or testing procedure.
2. **Systematic bias** - The laboratory's results are either consistently high or low for both samples when compared to the other participants (the plotted point falls near the top or bottom of the ellipse). This indicates that the participant is performing the test with a constant bias. Causes of systematic errors include improper calibration, the particular make/model of equipment or a modification to the testing procedure.
3. **Inconsistency in testing between samples/sample sets** - The laboratory's results indicate that there are differences in the way the two samples tested (the plotted point falls to the side of the ellipse). This type of error may be attributed to the analyst deviating from the procedure when testing one of the samples or a material interaction occurrence with the instrument or room conditions. The inconsistency is reflected in the CPVs for the two samples, such as a +1.5 CPV for sample A and a -2.2 CPV for sample B. CTS also will specify if the laboratory's data for one sample are high/low compared to the other participants. If this inconsistency is slight, the lab's plotted point will be an \* that falls on the edge of the ellipse.
4. **Inconsistency in testing within a sample** - The laboratory's within-lab standard deviation for a specified sample is high when compared to the other participants, often causing the lab's plotted point to fall outside of the ellipse.

---

Labs flagged with an \* are not typically included in the footnotes of a data table. These labs may locate their position in the control ellipse and use the definitions above to help identify the type of testing error. An \* should serve as a caution flag, a "yellow light", to a lab. If this error is repeated in future rounds, a lab may need to stop and review its testing procedures. The initial data flag is not cause for alarm. Interlaboratory tests conducted at regular intervals permit a lab to recognize trends in testing.



**Paper & Paperboard Interlaboratory Testing Program  
Analysis 350**

**Report #3152 G,  
December 2021**

**Color & Color Difference - Near White Papers - C/2deg obs  
Hunter L,a,b - Illuminant C - 2 Degree Observer**

Web Code	Data Flag	Samples	Hunter L, a, b Color Values			Color Difference Values				Instr Code
			L	a	b	$\Delta L$	$\Delta a$	$\Delta b$	$\Delta E$	
64EQCB		GA97	95.05	-0.67	4.09	-0.50	0.00	0.00	0.50	HE
		GA98	94.55	-0.67	4.09					
7C7ME6		GA97	94.39	-0.81	4.57	-0.71	-0.01	-0.04	0.71	VM
		GA98	93.68	-0.82	4.53					
CHD83B		GA97	94.24	-0.74	4.10	-0.48	0.00	0.08	0.49	LA
		GA98	93.76	-0.74	4.19					
CZV789		GA97	95.39	-0.89	4.16	-0.35	-0.01	0.06	0.35	LS
		GA98	95.04	-0.90	4.22					
JKCVED		GA97	95.17	-0.58	4.41	-0.38	-0.03	0.13	0.41	LS
		GA98	94.79	-0.61	4.53					
LZHWAQ	X	GA97	94.21	-0.81	4.02	-12.46	0.00	0.04	12.46X	TC
		GA98	81.74	-0.81	4.06					
NCLYLY		GA97	93.22	-0.41	3.66	-0.17	0.13	0.04	0.22	TS
		GA98	93.05	-0.28	3.70					
NRYBDR		GA97	93.69	-0.57	3.72	-0.60	0.08	0.11	0.61	TS
		GA98	93.10	-0.49	3.83					
P3DM6V		GA97	93.43	-0.52	3.49	-0.55	0.01	0.01	0.55	TS
		GA98	92.87	-0.52	3.50					
PXKMQZ		GA97	92.00	-0.36	2.50	0.26	-0.18	0.62	0.70	TS
		GA98	92.26	-0.54	3.12					
QX2Y7K		GA97	93.43	-0.45	3.70	-0.62	0.06	0.09	0.63	TS
		GA98	92.82	-0.39	3.79					
RW3KQ6		GA97	95.44	-0.83	3.95	-0.40	0.02	0.05	0.40	EH
		GA98	95.04	-0.81	4.01					
RWMWXQ		GA97	94.45	-0.74	4.23	-0.45	0.00	0.02	0.45	HE
		GA98	94.01	-0.74	4.25					
TNK9C4		GA97	93.99	-0.61	3.89	-0.49	-0.01	0.09	0.50	TC
		GA98	93.50	-0.63	3.98					
WWLPHR		GA97	94.41	-0.52	3.44	0.17	-0.06	0.33	0.38	XS
		GA98	94.58	-0.58	3.77					
YNKCC9		GA97	93.40	-0.28	3.54	-0.72	0.05	0.14	0.74	TS
		GA98	92.68	-0.22	3.69					



**Paper & Paperboard Interlaboratory Testing Program  
Analysis 350**

**Report #3152 G,  
December 2021**

**Color & Color Difference - Near White Papers - C/2deg obs  
Hunter L,a,b - Illuminant C - 2 Degree Observer**

Web Code	Data Flag	Samples	Hunter L, a, b Color Values			Color Difference Values				Instr Code
			L	a	b	$\Delta L$	$\Delta a$	$\Delta b$	$\Delta E$	
YQC9YG		GA97	94.78	-0.58	3.99	-0.42	-0.04	0.09	0.43	HE
		GA98	94.36	-0.62	4.08					
Z72MUZ		GA97	95.38	-0.83	4.01	-0.32	-0.02	0.05	0.33	TC
		GA98	95.06	-0.85	4.07					

Grand Means		Summary Statistics							
GA97	94.228	-0.622	3.860						
GA98	93.832	-0.623	3.966	-0.396	-0.001	0.110	0.493		
<u>Std Dev Btwn Labs</u>									
GA97	0.948	0.179	0.458						
GA98	0.927	0.193	0.348	0.269	0.065	0.154	0.146		

Statistics based on 17 of 18 reporting participants

**Comments on Assigned Data Flags for Test #350**

LZHWAQ (X) - Extreme data for "L" values for sample GA98. Very low delta "L" & very high delta "E" values.

**Key to Instrument Codes Reported by Participants**


EH	Datacolor Elrepho SF450	HE	Hunter LabScan
LA	L & W Elrepho AL300	LS	L & W Elrepho SE 070
TC	Technidyne Color Touch Series	TS	Technidyne Brightimeter Micro S-5
VM	Valmet PaperLab (was Kajaani/Robotest)	XS	X-Rite 938 Spectrodensitometer



**Paper & Paperboard Interlaboratory Testing Program**  
**Analysis 350**  
**Color & Color Difference - Near White Papers - C/2deg obs**  
**Hunter L,a,b - Illuminant C - 2 Degree Observer**

**Report #3152 G,**  
**December 2021**

Plot of L values GA98 vs L values GA97




If fewer than 20 laboratories are included in an analysis, a control ellipse will not be drawn on the two-sample plot.



**Paper & Paperboard Interlaboratory Testing Program**  
**Analysis 350**  
**Color & Color Difference - Near White Papers - C/2deg obs**  
**Hunter L,a,b - Illuminant C - 2 Degree Observer**

Report #3152 G,  
December 2021

Plot of a values GA98 vs a values GA97




If fewer than 20 laboratories are included in an analysis, a control ellipse will not be drawn on the two-sample plot.



**Paper & Paperboard Interlaboratory Testing Program**  
**Analysis 350**  
**Color & Color Difference - Near White Papers - C/2deg obs**  
**Hunter L,a,b - Illuminant C - 2 Degree Observer**

**Report #3152 G,**  
**December 2021**

Plot of b values GA98 vs b values GA97



If fewer than 20 laboratories are included in an analysis, a control ellipse will not be drawn on the two-sample plot.



**Paper & Paperboard Interlaboratory Testing Program  
Analysis 351**

**Report #3152 G,  
December 2021**

**Color & Color Difference - Near White Papers - D65/10deg obs  
Hunter L,a,b - Illuminant D65 - 10 Degree Observer**

Web Code	Data Flag	Samples	CIE L* a* b* Color Values			Color Difference Values				InstrCode
			L*	a*	b*	$\Delta L^*$	$\Delta a^*$	$\Delta b^*$	$\Delta E^*$	
2F6DUG		GA97	94.80	-0.74	3.17	0.12	0.00	0.60	0.61	NF
		GA98	94.92	-0.75	3.77					
66CZ8E		GA97	95.32	-0.61	4.14	-0.35	-0.02	0.12	0.37	EH
		GA98	94.97	-0.63	4.27					
6BEGCT		GA97	94.63	-0.46	3.86	-0.34	-0.05	-0.06	0.35	HE
		GA98	94.29	-0.51	3.80					
BHWQA4		GA97	95.34	-0.79	3.74	-0.29	-0.01	0.14	0.32	XC
		GA98	95.05	-0.80	3.88					
D9MQ6U		GA97	95.45	-0.54	4.51	-0.32	-0.03	0.03	0.33	NH
		GA98	95.13	-0.56	4.54					
EUEKB8		GA97	95.55	-0.55	4.07	-0.56	0.00	0.05	0.56	NG
		GA98	94.99	-0.55	4.12					
GXATV8		GA97	95.60	-0.61	4.14	-0.35	-0.03	0.02	0.35	HT
		GA98	95.25	-0.64	4.15					
H7ZZKQ		GA97	95.28	-0.55	4.07	-0.38	0.00	0.10	0.39	TC
		GA98	94.90	-0.55	4.17					
HR2UTG		GA97	95.08	-0.53	4.18	-0.56	-0.03	0.11	0.57	LS
		GA98	94.52	-0.56	4.29					
JKCVED		GA97	95.03	-0.83	4.40	-0.19	0.01	0.06	0.20	LS
		GA98	94.85	-0.82	4.46					
K43ARG		GA97	95.75	-0.56	4.11	-0.46	-0.01	0.10	0.48	XV
		GA98	95.29	-0.57	4.21					
KGXD4N		GA97	94.26	-0.49	3.70	0.79	-0.04	0.24	0.83 X	XB
		GA98	95.05	-0.53	3.94					
NKF6GL		GA97	92.69	-0.73	2.98	0.27	0.01	0.47	0.54	TC
		GA98	92.96	-0.72	3.45					
PYCQDF		GA97	95.44	-0.59	4.45	-0.36	-0.05	0.03	0.36	NG
		GA98	95.08	-0.64	4.48					
RW3KQ6		GA97	95.30	-0.62	4.11	-0.41	-0.03	0.09	0.42	EH
		GA98	94.89	-0.64	4.20					
TNK9C4		GA97	94.78	-0.66	4.17	-0.50	0.01	0.10	0.51	HE
		GA98	94.28	-0.65	4.26					



**Paper & Paperboard Interlaboratory Testing Program  
Analysis 351**

**Report #3152 G,  
December 2021**

**Color & Color Difference - Near White Papers - D65/10deg obs  
Hunter L,a,b - Illuminant D65 - 10 Degree Observer**

Z9KY9W	<b>GA97</b>	95.36	-0.69	4.16	-0.36	-0.04	0.03	0.36	HT
	<b>GA98</b>	95.00	-0.73	4.19					

<u>Grand Means</u>			<b>Summary Statistics</b>					
<b>GA97</b>	95.039	-0.620	3.996					
<b>GA98</b>	94.789	-0.639	4.128	-0.249	-0.018	0.132	0.443	
<u>Std Dev Btwn Labs</u>								
<b>GA97</b>	0.717	0.106	0.410					
<b>GA98</b>	0.553	0.096	0.283	0.343	0.019	0.166	0.148	
Statistics based on 17 of 17 reporting participants								

**Key to Instrument Codes Reported by Participants**


<b>EH</b> Datacolor Elrepho SF450	<b>HE</b> Hunter LabScan
<b>HT</b> Hunter UltraScan Vis	<b>LS</b> L & W Elrepho SE 070
<b>NF</b> Minolta CM-3600d Spectrophotometer	<b>NG</b> Minolta CM-3700d Spectrophotometer
<b>NH</b> Minolta CM-3700A Spectrophotometer	<b>TC</b> Technidyne Color Touch Series
<b>XB</b> X-Rite Ci7	<b>XC</b> X-Rite eXact Series
<b>XV</b> X-Rite SP60 Series	



**Paper & Paperboard Interlaboratory Testing Program**  
**Analysis 351**  
**Color & Color Difference - Near White Papers - D65/10deg obs**  
**Hunter L,a,b - Illuminant D65 - 10 Degree Observer**

Report #3152 G,  
December 2021

Plot of L values GA98 vs L values GA97




If fewer than 20 laboratories are included in an analysis, a control ellipse will not be drawn on the two-sample plot.



**Paper & Paperboard Interlaboratory Testing Program**  
**Analysis 351**  
**Color & Color Difference - Near White Papers - D65/10deg obs**  
**Hunter L,a,b - Illuminant D65 - 10 Degree Observer**

**Report #3152 G,**  
**December 2021**

Plot of a values GA98 vs a values GA97




If fewer than 20 laboratories are included in an analysis, a control ellipse will not be drawn on the two-sample plot.



**Paper & Paperboard Interlaboratory Testing Program**  
**Analysis 351**  
**Color & Color Difference - Near White Papers - D65/10deg obs**  
**Hunter L,a,b - Illuminant D65 - 10 Degree Observer**

Report #3152 G,  
December 2021

Plot of b values GA98 vs b values GA97



If fewer than 20 laboratories are included in an analysis, a control ellipse will not be drawn on the two-sample plot.



**Paper & Paperboard Interlaboratory Testing Program**  
**Analysis 360**  
**Thickness (Caliper), Printing papers**  
**TAPPI Official Test Method T411**

Report #3152G,  
December 2021

WebCode	Data Flag	Sample GV97			Sample GV98			Instr Code
		Lab Mean	Diff from Grand Mean	CPV	Lab Mean	Diff from Grand Mean	CPV	
2F6DUG		5.183	0.207	2.10	5.209	0.238	2.30	TM
2TET38		4.854	-0.123	-1.25	4.824	-0.147	-1.41	PP
2ZCK9F		5.011	0.034	0.34	4.978	0.007	0.07	TM
3MHNL9		4.976	-0.001	-0.01	4.945	-0.026	-0.25	TA
66CZ8E		4.863	-0.114	-1.16	4.854	-0.117	-1.12	EM
677PQA		5.047	0.070	0.71	4.978	0.007	0.07	TM
6BEGCT		5.055	0.078	0.79	5.013	0.042	0.41	PP
6QD44U		4.796	-0.181	-1.84	4.800	-0.171	-1.64	LW
8KXA2U		4.991	0.014	0.14	5.089	0.118	1.14	EM
8ZTP8F		5.090	0.113	1.15	5.096	0.125	1.21	TM
9DCURG		5.019	0.042	0.43	5.037	0.066	0.64	LW
9XVWMF		5.059	0.082	0.83	4.982	0.011	0.11	PP
9Z4W3V	*	4.958	-0.019	-0.19	4.821	-0.150	-1.44	LA
A4Y6LW	X	4.755	-0.222	-2.25	4.940	-0.031	-0.29	MT
AJVLBU	*	4.827	-0.150	-1.52	4.713	-0.258	-2.48	TA
AYP2JZ		5.008	0.031	0.32	4.954	-0.017	-0.16	PP
B2TW34		4.840	-0.137	-1.39	4.892	-0.079	-0.76	TM
BHWQA4		5.020	0.043	0.43	5.004	0.033	0.32	LW
C27VQK		5.115	0.138	1.40	5.131	0.160	1.54	LW
CHD83B		5.032	0.055	0.56	5.067	0.096	0.93	EM
CKEBRC		4.944	-0.032	-0.33	4.945	-0.025	-0.24	LW
CZV789		5.000	0.023	0.24	4.970	-0.001	-0.01	LW
D9MQ6U		5.132	0.155	1.58	5.126	0.155	1.50	PP
DMNHU6		5.042	0.065	0.66	5.084	0.113	1.09	LB
DQK4L9		5.050	0.073	0.74	5.063	0.092	0.89	LW
G74YQU		4.931	-0.046	-0.47	4.929	-0.042	-0.40	TA
GXATV8		4.993	0.016	0.16	4.971	0.000	0.00	EM
H7ZZKQ		4.976	-0.001	-0.01	4.976	0.006	0.06	PP
HCNDUR		4.887	-0.090	-0.91	4.970	-0.001	-0.01	LW
HWT43E		4.944	-0.033	-0.33	5.000	0.029	0.28	FR
JT7WZX		4.981	0.004	0.04	4.971	0.000	0.00	LW
KGXD4N		4.863	-0.114	-1.16	4.822	-0.149	-1.43	TM
LKRR3X		4.901	-0.076	-0.77	4.889	-0.082	-0.79	LA
LNPB4V		5.073	0.096	0.97	5.054	0.083	0.80	LW
NCLYLY		4.969	-0.008	-0.08	4.961	-0.010	-0.09	TM
NKF6GL	X	0.609	-4.368	-44.38	0.614	-4.357	-41.95	TA
NRYBDR		4.769	-0.208	-2.11	4.832	-0.139	-1.33	LA
P3DM6V	*	4.662	-0.315	-3.20	4.695	-0.276	-2.66	TM
PQLXBG		4.992	0.015	0.15	5.012	0.041	0.40	EM
PXKRYG		4.985	0.008	0.08	5.058	0.087	0.84	EM
PYCQDF		4.998	0.021	0.21	4.992	0.021	0.21	PP



**Paper & Paperboard Interlaboratory Testing Program**  
**Analysis 360**  
**Thickness (Caliper), Printing papers**  
**TAPPI Official Test Method T411**

**Report #3152G,**  
**December 2021**

WebCode	Data Flag	<u>Sample GV97</u>			<u>Sample GV98</u>			Instr Code
		Lab Mean	Diff from Grand Mean	CPV	Lab Mean	Diff from Grand Mean	CPV	
QX2Y7K		5.040	0.063	0.64	4.980	0.009	0.09	EM
RW3KQ6		5.041	0.064	0.65	4.982	0.011	0.11	EM
TD4HUL		4.995	0.018	0.18	5.006	0.035	0.34	TA
TQTEWK		5.026	0.049	0.50	5.013	0.043	0.41	LW
TYLRVJ		4.977	0.000	0.00	5.073	0.103	0.99	OK
WWLPHR		4.930	-0.047	-0.48	4.930	-0.041	-0.39	TM
Z72MUZ		5.081	0.104	1.06	4.997	0.026	0.25	TM
Z9KY9W		4.990	0.013	0.13	4.933	-0.038	-0.36	EM

<b>Summary Statistics</b>	<u>Sample GV97</u>	<u>Sample GV98</u>
<b>Grand Means</b>	4.98 mils	4.97 mils
<b>Std Dev Btwn Labs</b>	0.10 mils	0.10 mils

Statistics based on 47 of 49 reporting participants.

**Comments on Assigned Data Flags for Test #360**

NKF6GL (X) - Extreme Data.

A4Y6LW (X) - Inconsistent in testing between samples. Inconsistent within the determinations of sample GV98.

**Key to Instrument Codes Reported by Participants**

EM	Emveco	FR	Frank Instruments
LA	L & W Autoline	LB	L & W Autoline 600
LW	L & W	MT	Mitutoyo
OK	Oakland	PP	Technidyne Profile/Plus
TA	Thwing-Albert	TM	TMI





**Paper & Paperboard Interlaboratory Testing Program**  
**Analysis 361**  
**Thickness (Caliper), Packaging papers**  
**TAPPI Official Test Method T411**

Report #3152G,  
December 2021

WebCode	Data Flag	Sample GY97			Sample GY98			Instr Code
		Lab Mean	Diff from Grand Mean	CPV	Lab Mean	Diff from Grand Mean	CPV	
2QZN2N		7.762	0.179	1.33	7.849	0.267	1.93	LW
3MHNL9		7.629	0.046	0.34	7.601	0.019	0.14	TA
4FGQCB		7.598	0.015	0.11	7.611	0.029	0.21	LW
4JW9MZ		7.569	-0.014	-0.11	7.649	0.067	0.49	LW
4W48F8		7.798	0.215	1.60	7.673	0.091	0.66	LW
4WYPG9		7.623	0.040	0.30	7.622	0.040	0.29	LW
4Y6A69	*	7.287	-0.297	-2.22	7.213	-0.369	-2.66	LW
64EQCB		7.584	0.001	0.01	7.620	0.038	0.28	EM
66CZ8E		7.508	-0.075	-0.56	7.549	-0.033	-0.24	EM
6JUQ92		7.676	0.092	0.69	7.679	0.097	0.70	LA
7C7ME6		7.453	-0.130	-0.97	7.423	-0.159	-1.14	VP
7RHBJR		7.540	-0.043	-0.32	7.630	0.048	0.35	TA
946YRW		7.740	0.157	1.17	7.740	0.158	1.14	TM
AA6XFR		7.347	-0.237	-1.77	7.264	-0.318	-2.29	LA
AJVLBU	*	7.390	-0.193	-1.44	7.556	-0.026	-0.19	TA
CHD83B	*	7.859	0.276	2.06	7.713	0.131	0.94	EM
CKEBRC		7.570	-0.013	-0.10	7.576	-0.005	-0.04	LW
CM9YCT		7.668	0.085	0.63	7.677	0.095	0.69	EM
D9MQ6U		7.741	0.158	1.18	7.804	0.222	1.60	PP
DMNHU6		7.707	0.124	0.92	7.707	0.125	0.90	LB
E38ZUJ		7.580	-0.003	-0.02	7.596	0.014	0.10	LA
EJ33VX		7.680	0.097	0.72	7.682	0.100	0.72	PP
ERTN2V		7.583	0.000	0.00	7.575	-0.006	-0.05	LW
HCNDUR		7.587	0.003	0.03	7.551	-0.031	-0.22	LW
HGKCXP		7.417	-0.166	-1.24	7.366	-0.216	-1.55	GE
HR2UTG		7.586	0.003	0.02	7.569	-0.013	-0.09	LW
JBQRD7		7.559	-0.024	-0.18	7.533	-0.049	-0.35	TM
JKCVED		7.264	-0.319	-2.39	7.307	-0.275	-1.98	TM
K3UAC3		7.693	0.110	0.82	7.676	0.094	0.68	TM
PXKMQZ		7.526	-0.057	-0.43	7.465	-0.117	-0.84	OK
RWMWXQ		7.608	0.025	0.19	7.524	-0.058	-0.42	EM
TNK9C4		7.557	-0.026	-0.20	7.619	0.037	0.27	EM
W83JPD		7.587	0.003	0.03	7.654	0.072	0.52	LW
WQJNB6	X	9.522	1.939	14.49	10.145	2.563	18.49	LW
WYUZCA		7.508	-0.075	-0.56	7.512	-0.070	-0.50	LA
YQC9YG		7.631	0.048	0.36	7.575	-0.006	-0.05	EM



**Paper & Paperboard Interlaboratory Testing Program**

**Report #3152G,  
December 2021**

**Analysis 361**

**Thickness (Caliper), Packaging papers**

**TAPPI Official Test Method T411**

<b>Summary Statistics</b>	<b>Sample GY97</b>	<b>Sample GY98</b>
<b>Grand Means</b>	7.58 mils	7.58 mils
<b>Stnd Dev Btwn Labs</b>	0.13 mils	0.14 mils

Statistics based on 35 of 36 reporting participants.

**Comments on Assigned Data Flags for Test #361**

WQJNB6 (X) - Extreme Data.

**Key to Instrument Codes Reported by Participants**

<b>EM</b>	Emveco	<b>GE</b>	Gester Electronic Thickness Tester
<b>LA</b>	L & W Autoline	<b>LB</b>	L & W Autoline 600
<b>LW</b>	L & W	<b>OK</b>	Oakland
<b>PP</b>	Technidyne Profile/Plus	<b>TA</b>	Thwing-Albert
<b>TM</b>	TMI	<b>VP</b>	Valmet Paper Lab Automated Tester



# Paper & Paperboard Interlaboratory Testing Program

Report #3152G,  
December 2021

## Analysis 361


Thickness (Caliper), Packaging papers

TAPPI Official Test Method T411

Grand Mean Sample GY97 = 7.5832  
mils

Grand Mean Sample GY98 = 7.5817  
mils

ANALYSIS 361





**Paper & Paperboard Interlaboratory Testing Program**  
**Analysis 364**  
**Coefficient of Static Friction - Horizontal Plane Method - Printing Papers**  
**TAPPI Official Test Method T549**

**Report #3152G,**  
**December 2021**

WebCode	Data Flag	<u>Sample GD97</u>			<u>Sample GD98</u>			Instr Code
		Lab Mean	Diff from Grand Mean	CPV	Lab Mean	Diff from Grand Mean	CPV	
2TET38		0.4904	-0.0768	-1.08	0.4564	-0.1142	-1.69	TA
4W48F8		0.6066	0.0394	0.56	0.6154	0.0448	0.66	TA
AA6XFR		0.5556	-0.0116	-0.16	0.5546	-0.0160	-0.24	TA
D9MQ6U		0.6720	0.1048	1.48	0.6360	0.0654	0.97	TP
KKWUEW		0.6034	0.0362	0.51	0.6066	0.0360	0.53	TA
PQLXBG		0.5700	0.0028	0.04	0.5840	0.0134	0.20	TA
QX2Y7K		0.6214	0.0542	0.77	0.6564	0.0858	1.27	TA
VBQL4E		0.6162	0.0490	0.69	0.6054	0.0348	0.51	IT
WWLPHR		0.4782	-0.0890	-1.26	0.4842	-0.0864	-1.28	XX
YNKCC9		0.4580	-0.1092	-1.54	0.5070	-0.0636	-0.94	TA

<b>Summary Statistics</b>	<u>Sample GD97</u>	<u>Sample GD98</u>
<b>Grand Means</b>	0.57 COF	0.57 COF
<b>Std Dev Btwn Labs</b>	0.07 COF	0.07 COF
Statistics based on 10 of 10 reporting participants.		

**Key to Instrument Codes Reported by Participants**

IT	IMASS SP-2100	TA	Thwing-Albert Friction Tester
TP	TMI 32-25 COF Tester (Inclined Plane)	XX	Instrument make/model not specified by lab




**Paper & Paperboard Interlaboratory Testing Program**  
**Analysis 364**  
**Coefficient of Static Friction - Horizontal Plane Method - Printing Papers**  
**TAPPI Official Test Method T549**

**Report #3152G,**  
**December 2021**

**Grand Mean Sample GD97 = 0.56718**  
**COF**

**Grand Mean Sample GD98 =**  
**0.57060 COF**



If fewer than 20 laboratories are included in an analysis, a control ellipse will not be drawn on the two-sample plot.



**Paper & Paperboard Interlaboratory Testing Program**  
**Analysis 365**  
**Coefficient of Kinetic Friction - Horizontal Plane Method - Printing Papers**  
**TAPPI Official Test Method T549**

**Report #3152G,**  
**December 2021**

WebCode	Data Flag	<u>Sample GD97</u>			<u>Sample GD98</u>			Instr Code
		Lab Mean	Diff from Grand Mean	CPV	Lab Mean	Diff from Grand Mean	CPV	
2TET38		0.3934	-0.0589	-0.87	0.3824	-0.0720	-1.14	TA
4W48F8		0.5238	0.0715	1.06	0.5088	0.0544	0.86	TN
AA6XFR		0.4928	0.0405	0.60	0.4936	0.0392	0.62	TA
KKWUEW		0.4628	0.0105	0.15	0.4886	0.0342	0.54	TA
PQLXBG		0.4700	0.0177	0.26	0.4900	0.0356	0.56	XX
QX2Y7K		0.5084	0.0561	0.83	0.5344	0.0800	1.26	TA
VBQL4E		0.4512	-0.0011	-0.02	0.3946	-0.0598	-0.94	IR
WWLPHR		0.4668	0.0145	0.21	0.4426	-0.0118	-0.19	XX
YNKCC9		0.3016	-0.1507	-2.22	0.3548	-0.0996	-1.57	TA

<b>Summary Statistics</b>	<u>Sample GD97</u>	<u>Sample GD98</u>
<b>Grand Means</b>	0.45 COF	0.45 COF
<b>Stnd Dev Btwn Labs</b>	0.07 COF	0.06 COF

Statistics based on 9 of 9 reporting participants.

**Key to Instrument Codes Reported by Participants**

IR	IMASS SP-2000	TA	Thwing-Albert Friction Tester
TN	TMI 32-07 Monitor/Slip and Friction	XX	Instrument make/model not specified by lab




**Paper & Paperboard Interlaboratory Testing Program**  
**Analysis 365**  
**Coefficient of Kinetic Friction - Horizontal Plane Method - Printing Papers**  
**TAPPI Official Test Method T549**

**Report #3152G,**  
**December 2021**

**Grand Mean Sample GD97 = 0.45231**  
**COF**

**Grand Mean Sample GD98 =**  
**0.45442 COF**



If fewer than 20 laboratories are included in an analysis, a control ellipse will not be drawn on the two-sample plot.



# Paper & Paperboard Interlaboratory Testing Program

Report #3152G,  
December 2021

## Analysis 370

### Air Resistance - Gurley Oil Type

### TAPPI Official Test Method T460

WebCode	Data Flag	Sample GE97			Sample GE98			Instr Code
		Lab Mean	Diff from Grand Mean	CPV	Lab Mean	Diff from Grand Mean	CPV	
2F6DUG		13.15	-0.64	-0.94	13.31	-0.50	-0.67	LP
3MHNL9		13.64	-0.15	-0.22	13.65	-0.16	-0.21	PP
4FGQCB		13.87	0.08	0.12	14.32	0.51	0.70	LW
4JW9MZ		14.44	0.65	0.95	14.65	0.84	1.15	LP
4WYPG9		12.77	-1.02	-1.49	13.13	-0.68	-0.92	HM
6BEGCT		13.76	-0.03	-0.04	13.78	-0.03	-0.04	PP
772P7Z		14.75	0.96	1.40	14.76	0.95	1.30	TM
7C7ME6		14.92	1.13	1.65	14.72	0.91	1.24	VM
7WAFCW		12.86	-0.93	-1.36	12.48	-1.33	-1.80	LA
8KXA2U		13.67	-0.12	-0.18	13.58	-0.23	-0.31	PP
8ZTP8F		14.09	0.30	0.44	14.14	0.33	0.45	HG
9XVWMF		13.59	-0.20	-0.30	13.81	0.00	0.00	PP
AA6XFR		13.87	0.08	0.12	14.05	0.24	0.33	LA
AYP2JZ		14.57	0.78	1.13	14.97	1.16	1.58	PP
BHWQA4		14.30	0.51	0.74	14.10	0.29	0.40	LW
C27VQK		13.70	-0.09	-0.13	13.24	-0.57	-0.77	LP
CM9YCT		13.95	0.16	0.23	14.21	0.40	0.55	LP
CZV789		13.32	-0.47	-0.69	13.67	-0.14	-0.19	LP
D9MQ6U		12.97	-0.83	-1.21	13.07	-0.74	-1.01	PP
DQK4L9		14.01	0.22	0.32	14.17	0.36	0.50	LP
G74YQU		13.66	-0.13	-0.19	13.90	0.09	0.13	GA
GXATV8		13.84	0.05	0.07	13.96	0.15	0.21	PP
H7ZZKQ		14.32	0.53	0.78	14.71	0.90	1.22	PP
HCNDUR		13.56	-0.23	-0.34	13.36	-0.45	-0.61	PP
HTEGBV		12.31	-1.48	-2.16	12.29	-1.52	-2.06	GA
KGXD4N		13.89	0.09	0.14	13.67	-0.14	-0.19	PP
KKWUEW		14.45	0.66	0.96	14.56	0.75	1.03	WG
LKRR3X	*	15.94	2.15	3.14	15.80	1.99	2.71	LA
M4XQHB		14.41	0.62	0.90	13.92	0.11	0.16	XX
MKLYML		14.25	0.46	0.67	13.79	-0.02	-0.02	GL
NC99XN	*	13.10	-0.69	-1.01	12.37	-1.44	-1.95	LP
NCLYLY		12.40	-1.39	-2.03	12.23	-1.58	-2.14	LW
NKF6GL		13.16	-0.64	-0.93	13.54	-0.27	-0.37	PP
NRYBDR		13.47	-0.32	-0.47	13.89	0.08	0.11	LA
NUQ4DH		13.46	-0.33	-0.48	13.26	-0.55	-0.74	LP
P3DM6V		14.42	0.63	0.92	14.10	0.29	0.40	LP
PQLXBG		13.65	-0.14	-0.20	13.67	-0.14	-0.18	PP
RW3KQ6		13.73	-0.06	-0.09	13.37	-0.43	-0.59	PP
TK6UQJ		14.08	0.29	0.42	14.30	0.49	0.67	TL
TNK9C4		14.03	0.24	0.35	14.27	0.47	0.63	PP
W83JPD		12.92	-0.87	-1.27	12.84	-0.97	-1.31	LP



**Paper & Paperboard Interlaboratory Testing Program**

**Report #3152G,  
December 2021**

**Analysis 370**

**Air Resistance - Gurley Oil Type**

**TAPPI Official Test Method T460**

WebCode	Data Flag	Sample GE97			Sample GE98			Instr Code
		Lab Mean	Diff from Grand Mean	CPV	Lab Mean	Diff from Grand Mean	CPV	
WQJNB6		13.14	-0.65	-0.95	13.07	-0.74	-1.01	TL
WWLPHR		14.60	0.81	1.18	14.40	0.59	0.81	GS
WYUZCA		13.92	0.13	0.19	14.37	0.56	0.77	LA
Z9KY9W		13.70	-0.09	-0.13	13.85	0.04	0.06	HG

Summary Statistics	Sample GE97	Sample GE98
<b>Grand Means</b>	13.79 sec/100 cc	13.81 sec/100 cc
<b>Std Dev Btwn Labs</b>	0.68 sec/100 cc	0.74 sec/100 cc
Statistics based on 45 of 45 reporting participants.		

**Key to Instrument Codes Reported by Participants**

<b>GA</b> Gurley Precision #4340 Automatic Densometer	<b>GL</b> Gurley #4110
<b>GS</b> Gurley-Hill S-P-S Tester #4190	<b>HG</b> Technidyne - Hagerty Model #1
<b>HM</b> Technidyne - Hagerty Model #538	<b>LA</b> L & W Autoline
<b>LP</b> L & W Densometer, Air Permeance	<b>LW</b> L & W Type Gurley Densometer, Oil Flotation
<b>PP</b> Technidyne Profile/Plus	<b>TL</b> Gurley Densometer #4110, Oil Flotation
<b>TM</b> TMI Densometer 58-03	<b>VM</b> Valmet PaperLab (was Kajaani/Robotest)
<b>WG</b> W & LE Gurley Tester	<b>XX</b> Instrument make/model not specified by lab




**Paper & Paperboard Interlaboratory Testing Program**  
**Analysis 370**  
**Air Resistance - Gurley Oil Type**  
**TAPPI Official Test Method T460**

**Report #3152G,**  
**December 2021**

**Grand Mean Sample GE97 = 13.791**  
sec/100 cc

**Grand Mean Sample GE98 = 13.806**  
sec/100 cc





**Paper & Paperboard Interlaboratory Testing Program**  
**Analysis 372**  
**Porosity - Sheffield Type - Sheffield Units for 3/4 inch Diameter Orifice**  
**TAPPI Official Test Method T547**

**Report #3152G,**  
**December 2021**

WebCode	Data Flag	<u>Sample GE97</u>			<u>Sample GE98</u>			Instr Code
		Lab Mean	Diff from Grand Mean	CPV	Lab Mean	Diff from Grand Mean	CPV	
3MHNL9		187.4	-1.2	-0.10	188.4	-3.9	-0.27	PP
677PQA	X	256.4	67.8	5.61	261.6	69.3	4.87	LP
7C7ME6		190.8	2.2	0.18	192.4	0.1	0.01	PP
NKF6GL		192.9	4.3	0.36	205.4	13.1	0.92	TT
TD4HUL		191.3	2.7	0.22	194.3	2.0	0.14	HM
TYLRVJ		202.8	14.2	1.18	206.0	13.7	0.96	LA
WWLPHR		166.3	-22.3	-1.84	167.1	-25.2	-1.77	SH

<b>Summary Statistics</b>	<u>Sample GE97</u>	<u>Sample GE98</u>
<b>Grand Means</b>	188.58 Sheffield Units	192.27 Sheffield Units
<b>Std Dev Btwn Labs</b>	12.09 Sheffield Units	14.24 Sheffield Units
Statistics based on 6 of 7 reporting participants.		

**Comments on Assigned Data Flags for Test #372**

677PQA (X) - Data for both samples are high. Possible Systematic Error.

**Key to Instrument Codes Reported by Participants**

HM	Technidyne - Hagerty Model #538	LA	L & W Roughness Sheffield - Autoline
LP	L & W Densometer, Air Permeance	PP	Technidyne Profile/Plus
SH	Sheffield	TT	TMI Monitor/Smoothness II, Model 58-24




**Paper & Paperboard Interlaboratory Testing Program**  
**Analysis 372**  
**Porosity - Sheffield Type - Sheffield Units for 3/4 inch Diameter Orifice**  
**TAPPI Official Test Method T547**

**Report #3152G,**  
**December 2021**

**Grand Mean Sample GE97 = 188.58**  
**Sheffield Units**

**Grand Mean Sample GE98 = 192.27**  
**Sheffield Units**

**ANALYSIS 372**



If fewer than 20 laboratories are included in an analysis, a control ellipse will not be drawn on the two-sample plot.



**Paper & Paperboard Interlaboratory Testing Program**

**Report #3152G,  
December 2021**

**Analysis 376**

**Roughness - Print Surf Method - 0.5 to 4.0 Microns**

**TAPPI Official Test Method T555**

WebCode	Data Flag	Sample GJ97			Sample GJ98			Instr Code
		Lab Mean	Diff from Grand Mean	CPV	Lab Mean	Diff from Grand Mean	CPV	
64EQCB		0.9260	-0.0168	-0.17	0.9670	0.0196	0.19	ZZ
66CZ8E		1.0150	0.0722	0.72	0.9830	0.0356	0.34	ZZ
6BEGCT		0.8930	-0.0498	-0.50	0.9860	0.0386	0.37	ZZ
7C7ME6	X	1.1240	0.1812	1.80	0.9500	0.0026	0.03	ZZ
8KXA2U		1.0890	0.1462	1.45	1.0920	0.1446	1.39	ZZ
93BB92		1.0670	0.1242	1.24	1.1140	0.1666	1.60	ZZ
9DCURG		1.0110	0.0682	0.68	0.9980	0.0506	0.49	ZZ
9Z4W3V		0.8090	-0.1338	-1.33	0.8580	-0.0894	-0.86	ZZ
CKEBRC		0.8390	-0.1038	-1.03	0.8310	-0.1164	-1.12	ZZ
DMNHU6		0.9170	-0.0258	-0.26	0.8850	-0.0624	-0.60	ZZ
E38ZUJ		0.7510	-0.1918	-1.91	0.7540	-0.1934	-1.85	ZZ
HR2UTG		0.9070	-0.0358	-0.36	0.8700	-0.0774	-0.74	ZZ
JKCVED		0.9470	0.0042	0.04	0.9460	-0.0014	-0.01	ZZ
KKWUEW		0.8450	-0.0978	-0.97	0.7850	-0.1624	-1.56	ZZ
KP2XVH		0.8530	-0.0898	-0.89	0.8900	-0.0574	-0.55	ZZ
L7Q34C		0.7700	-0.1728	-1.72	0.7680	-0.1794	-1.72	ZZ
LZHWAQ		0.9800	0.0372	0.37	0.9470	-0.0004	0.00	ZZ
NKF6GL		0.9990	0.0562	0.56	1.0720	0.1246	1.20	ZZ
P3DM6V		0.9970	0.0542	0.54	0.9640	0.0166	0.16	ZZ
PXKMQZ		0.9530	0.0102	0.10	0.9160	-0.0314	-0.30	ZZ
RW3KQ6	*	1.2060	0.2632	2.62	1.1820	0.2346	2.25	ZZ
RWMWXQ		0.9530	0.0102	0.10	1.0060	0.0586	0.56	ZZ
TNK9C4		0.8830	-0.0598	-0.60	0.9040	-0.0434	-0.42	ZZ
TQTEWK		0.9070	-0.0358	-0.36	0.8600	-0.0874	-0.84	ZZ
VQ46R9		0.9620	0.0192	0.19	0.9860	0.0386	0.37	ZZ
YNKCC9		0.9610	0.0182	0.18	0.9540	0.0066	0.06	ZZ
YQC9YG		0.9120	-0.0308	-0.31	0.9350	-0.0124	-0.12	ZZ
YVMB2C		1.1000	0.1572	1.56	1.0920	0.1446	1.39	ZZ
Z72MUZ		0.9470	0.0042	0.04	0.9810	0.0336	0.32	ZZ

Summary Statistics	Sample GJ97	Sample GJ98
<b>Grand Means</b>	0.94 Microns	0.95 Microns
<b>Std Dev Btwn Labs</b>	0.10 Microns	0.10 Microns

Statistics based on 28 of 29 reporting participants.

**Comments on Assigned Data Flags for Test #376**

7C7ME6 (X) - Inconsistent in testing between samples.



**Paper & Paperboard Interlaboratory Testing Program**

**Report #3152G,  
December 2021**

**Analysis 376**

**Roughness - Print Surf Method - 0.5 to 4.0 Microns**

**TAPPI Official Test Method T555**

**Key to Instrument Codes Reported by Participants**

**ZZ** Instruments No Longer Tracked



# Paper & Paperboard Interlaboratory Testing Program

Report #3152G,  
December 2021


## Analysis 376

Roughness - Print Surf Method - 0.5 to 4.0 Microns

TAPPI Official Test Method T555

Grand Mean Sample GJ97 = 0.94282  
Microns

Grand Mean Sample GJ98 =  
0.94736 Microns





**Paper & Paperboard Interlaboratory Testing Program**  
**Analysis 377**  
**Roughness - Print Surf Method - 2.5 to 6.0 Microns**  
**TAPPI Official Test Method T555**

**Report #3152G,**  
**December 2021**

WebCode	Data Flag	<u>Sample GK97</u>			<u>Sample GK98</u>			Instr Code
		Lab Mean	Diff from Grand Mean	CPV	Lab Mean	Diff from Grand Mean	CPV	
4W48F8		5.959	0.159	0.88	5.775	-0.071	-0.30	ZZ
66CZ8E		5.667	-0.133	-0.73	5.578	-0.268	-1.12	ZZ
D9MQ6U		5.774	-0.026	-0.14	5.944	0.098	0.41	ZZ
DMNHU6		5.808	0.008	0.05	5.849	0.003	0.01	ZZ
HCNDUR		5.812	0.012	0.07	5.595	-0.251	-1.05	ZZ
KKWUEW		5.622	-0.178	-0.98	5.618	-0.228	-0.95	ZZ
NRYBDR		5.484	-0.316	-1.75	5.693	-0.153	-0.64	ZZ
PQLXBG		5.940	0.140	0.78	5.972	0.126	0.53	ZZ
QX2Y7K		5.816	0.016	0.09	6.212	0.366	1.53	ZZ
TNK9C4		6.115	0.315	1.74	6.222	0.376	1.57	ZZ

<b>Summary Statistics</b>	<u>Sample GK97</u>	<u>Sample GK98</u>
<b>Grand Means</b>	5.80 Microns	5.85 Microns
<b>Std Dev Btwn Labs</b>	0.18 Microns	0.24 Microns
Statistics based on 10 of 10 reporting participants.		

**Key to Instrument Codes Reported by Participants**

ZZ Instruments No Longer Tracked




**Paper & Paperboard Interlaboratory Testing Program**  
**Analysis 377**  
**Roughness - Print Surf Method - 2.5 to 6.0 Microns**  
**TAPPI Official Test Method T555**

**Report #3152G,**  
**December 2021**

**Grand Mean Sample GK97 = 5.7997**  
**Microns**

**Grand Mean Sample GK98 = 5.8458**  
**Microns**



If fewer than 20 laboratories are included in an analysis, a control ellipse will not be drawn on the two-sample plot.



**Paper & Paperboard Interlaboratory Testing Program**  
**Analysis 378**  
**Roughness - Sheffield Type**  
**TAPPI Official Test Method T538**

Report #3152G,  
December 2021

WebCode	Data Flag	Sample GL97			Sample GL98			Instr Code
		Lab Mean	Diff from Grand Mean	CPV	Lab Mean	Diff from Grand Mean	CPV	
3MHNL9		120.3	-3.3	-0.43	125.2	0.5	0.05	PP
4DM3TE		123.5	-0.1	-0.01	113.8	-10.9	-1.08	LA
4JW9MZ		125.3	1.7	0.23	124.3	-0.4	-0.04	LW
4W48F8		125.3	1.7	0.23	127.6	2.9	0.28	LW
64EQCB		118.7	-4.9	-0.65	112.2	-12.5	-1.23	PP
66CZ8E		125.1	1.5	0.20	130.5	5.8	0.57	LW
677PQA		120.2	-3.4	-0.45	118.4	-6.3	-0.62	LW
6BEGCT		119.5	-4.1	-0.54	119.8	-5.0	-0.49	PP
7C7ME6		127.0	3.4	0.45	115.1	-9.6	-0.95	VM
7WAFCW		121.3	-2.3	-0.30	117.0	-7.7	-0.76	LA
8KXA2U		116.6	-7.0	-0.93	123.9	-0.9	-0.08	PP
8ZTP8F		130.8	7.2	0.95	130.6	5.9	0.58	TS
93BB92		127.7	4.1	0.54	124.9	0.2	0.02	LW
9XVWMF		113.2	-10.3	-1.37	114.6	-10.1	-1.00	PP
AYP2JZ		130.6	7.0	0.93	127.7	3.0	0.29	PP
BHWQA4		126.7	3.1	0.41	123.9	-0.8	-0.08	TS
CZV789	*	128.8	5.2	0.69	151.8	27.1	2.67	LW
D9MQ6U		133.3	9.7	1.29	143.8	19.0	1.88	PP
DMNHU6		123.7	0.1	0.02	128.2	3.5	0.34	LB
E38ZUJ		125.6	2.0	0.27	129.9	5.2	0.51	LA
G74YQU		126.6	3.0	0.40	114.1	-10.7	-1.05	GA
GUCGTF		135.8	12.2	1.62	138.0	13.3	1.31	TT
GXATV8		120.9	-2.7	-0.35	120.4	-4.3	-0.43	SH
H7ZZKQ		117.6	-6.0	-0.79	126.5	1.8	0.18	PP
HCNDUR		125.2	1.6	0.21	126.6	1.9	0.18	PP
HR2UTG		126.2	2.6	0.34	131.4	6.7	0.66	PP
JKCVED	*	138.6	15.0	1.99	150.4	25.7	2.54	TT
KGXD4N		113.7	-9.8	-1.30	115.3	-9.4	-0.93	PP
KKWUEW	*	143.9	20.3	2.69	145.0	20.3	2.00	XX
L7Q34C		121.2	-2.4	-0.32	114.6	-10.1	-1.00	LW
LKRR3X		107.4	-16.2	-2.14	115.0	-9.7	-0.96	LA
NCLYLY		127.9	4.3	0.57	129.5	4.8	0.47	SH
NKF6GL		135.5	11.9	1.58	126.8	2.1	0.21	TT
NRYBDR		124.6	1.0	0.13	128.8	4.1	0.40	LA
P3DM6V		118.8	-4.8	-0.63	121.1	-3.6	-0.36	TS
PQLXBG		121.2	-2.3	-0.31	123.0	-1.8	-0.17	PP
PXKMQZ	X	167.5	43.9	5.81	174.0	49.3	4.87	GL
PYCQDF		123.1	-0.5	-0.06	113.1	-11.6	-1.15	PP
QX2Y7K		121.2	-2.4	-0.32	119.2	-5.5	-0.54	PP
RJCL6U		123.3	-0.3	-0.04	129.4	4.7	0.46	GA
RWMWXQ		123.3	-0.3	-0.03	129.7	5.0	0.49	PP



**Paper & Paperboard Interlaboratory Testing Program**  
**Analysis 378**  
**Roughness - Sheffield Type**  
**TAPPI Official Test Method T538**

**Report #3152G,**  
**December 2021**

WebCode	Data Flag	Sample GL97			Sample GL98			Instr Code
		Lab Mean	Diff from Grand Mean	CPV	Lab Mean	Diff from Grand Mean	CPV	
TNK9C4		124.3	0.7	0.09	133.0	8.3	0.82	LW
TYLRVJ		112.6	-11.0	-1.45	124.4	-0.3	-0.03	LA
WB4MGZ	*	132.5	8.9	1.18	111.8	-12.9	-1.28	TT
WWLPHR		122.1	-1.5	-0.20	114.7	-10.0	-0.99	XX
YJRB8M		105.1	-18.5	-2.44	106.6	-18.1	-1.79	MP
YNKCC9		121.8	-1.8	-0.24	133.8	9.1	0.90	HM
YQC9YG		125.2	1.6	0.22	135.9	11.2	1.10	PP
YVMB2C		106.8	-16.8	-2.22	109.6	-15.1	-1.49	LA
Z9KY9W		125.9	2.3	0.31	120.2	-4.5	-0.45	HM

Summary Statistics	Sample GL97	Sample GL98
<b>Grand Means</b>	123.58 Sheffield	124.71 Sheffield
<b>Std Dev Btwn Labs</b>	7.56 Sheffield	10.13 Sheffield

Statistics based on 49 of 50 reporting participants.

**Comments on Assigned Data Flags for Test #378**

PXKMQZ (X) - Data for both samples are high.

**Key to Instrument Codes Reported by Participants**

<b>GA</b> Gurley Precision #4340 Automatic Densometer	<b>GL</b> Giddings and Lewis Sheffield
<b>HM</b> Technidyne - Hagerty Model #538	<b>LA</b> L & W Roughness Sheffield - Autoline
<b>LB</b> L & W - Autoline 600	<b>LW</b> L & W Roughness Tester
<b>MP</b> Metso Paperlab	<b>PP</b> Technidyne Profile/Plus
<b>SH</b> Sheffield (Bendix Precisionaire)	<b>TS</b> TMI Monitor/Smoothness, Model 58-02
<b>TT</b> TMI Monitor/Smoothness II, Model 58-24	<b>VM</b> Valmet PaperLab (was Kajaani\Robotest)
<b>XX</b> Instrument make/model not specified by lab	



# Paper & Paperboard Interlaboratory Testing Program

Report #3152G,  
December 2021

## Analysis 378


### Roughness - Sheffield Type

#### TAPPI Official Test Method T538

Grand Mean Sample GL97 = 123.58  
Sheffield

Grand Mean Sample GL98 = 124.71  
Sheffield

ANALYSIS 378





**Paper & Paperboard Interlaboratory Testing Program**  
**Analysis 382**  
**Moisture in Paper**  
**TAPPI Official Test Method T412**

Report #3152G,  
December 2021

WebCode	Data Flag	Sample GM97			Sample GM98			Instr Code
		Lab Mean	Diff from Grand Mean	CPV	Lab Mean	Diff from Grand Mean	CPV	
2KDHM2		4.685	0.343	0.83	3.808	-0.528	-1.22	ZZ
2W76TW		4.761	0.419	1.02	4.634	0.298	0.69	ZZ
2YXZF7		4.352	0.010	0.02	4.860	0.524	1.21	ZZ
6FVFD2		4.870	0.528	1.28	4.670	0.334	0.77	ZZ
9DCURG		3.979	-0.363	-0.88	4.001	-0.335	-0.78	ZZ
CRWNNQ		3.970	-0.372	-0.90	4.620	0.284	0.66	ZZ
CZV789		3.871	-0.471	-1.15	3.832	-0.504	-1.17	ZZ
EUEKB8		3.730	-0.612	-1.49	3.800	-0.536	-1.24	ZZ
GHXCN7		3.925	-0.417	-1.01	3.910	-0.426	-0.99	ZZ
HGKCXP		4.359	0.017	0.04	4.158	-0.178	-0.41	ZZ
JKCVED		4.872	0.530	1.29	5.010	0.674	1.56	ZZ
K3UAC3		4.214	-0.128	-0.31	4.171	-0.165	-0.38	ZZ
PQLXBG		4.321	-0.021	-0.05	4.406	0.070	0.16	ZZ
TD4HUL		4.881	0.539	1.31	4.825	0.489	1.13	ZZ

Summary Statistics	Sample GM97	Sample GM98
<b>Grand Means</b>	4.34 Percent	4.34 Percent
<b>Std Dev Btwn Labs</b>	0.41 Percent	0.43 Percent

Statistics based on 14 of 14 reporting participants.

**Key to Instrument Codes Reported by Participants**

ZZ Instruments No Longer Tracked



# Paper & Paperboard Interlaboratory Testing Program


Report #3152G,  
December 2021

## Analysis 382 Moisture in Paper

### TAPPI Official Test Method T412

Grand Mean Sample GM97 = 4.3421  
Percent

Grand Mean Sample GM98 = 4.3361  
Percent



If fewer than 20 laboratories are included in an analysis, a control ellipse will not be drawn on the two-sample plot.



**Paper & Paperboard Interlaboratory Testing Program**

**Report #3152G,  
December 2021**

**Analysis 384**

**Opacity (89% Reflectance Backing) - Fine Papers**

**TAPPI Official Test Method T425**

WebCode	Data Flag	<u>Sample GN97</u>			<u>Sample GN98</u>			Instr Code
		Lab Mean	Diff from Grand Mean	CPV	Lab Mean	Diff from Grand Mean	CPV	
2F6DUG		93.73	0.32	0.19	93.68	0.17	0.10	ZZ
3MHNL9		93.17	-0.24	-0.14	93.44	-0.07	-0.04	ZZ
6BEGCT		94.63	1.22	0.70	95.31	1.80	1.08	ZZ
8ZTP8F		93.60	0.19	0.11	93.51	0.00	0.00	ZZ
9Z4W3V		97.93	4.52	2.58	97.81	4.30	2.59	ZZ
AYP2JZ		93.70	0.29	0.17	93.74	0.23	0.14	ZZ
BHWQA4		93.06	-0.35	-0.20	93.24	-0.27	-0.16	ZZ
D9MQ6U		92.28	-1.12	-0.64	93.50	-0.01	0.00	ZZ
GXATV8		93.59	0.18	0.10	93.65	0.14	0.09	ZZ
H7ZZKQ	*	93.66	0.25	0.14	92.47	-1.04	-0.63	ZZ
HCNDUR		93.58	0.18	0.10	93.44	-0.07	-0.04	ZZ
KGXD4N		93.16	-0.25	-0.14	93.73	0.22	0.13	ZZ
LKRR3X		88.96	-4.45	-2.54	89.38	-4.13	-2.48	ZZ
LZHWAQ		93.91	0.50	0.29	93.81	0.30	0.18	ZZ
NKF6GL		94.03	0.62	0.36	93.53	0.02	0.01	ZZ
NRYBDR		94.27	0.87	0.49	94.51	1.00	0.60	ZZ
PQLXBG		93.57	0.16	0.09	94.27	0.76	0.46	ZZ
PYQDF		93.69	0.28	0.16	94.11	0.60	0.36	ZZ
QX2Y7K		93.72	0.31	0.18	93.35	-0.15	-0.09	ZZ
R6TXVW		93.04	-0.36	-0.21	93.31	-0.20	-0.12	ZZ
RW3KQ6		93.55	0.14	0.08	93.70	0.19	0.12	ZZ
TYLRVJ		93.44	0.03	0.02	93.40	-0.11	-0.06	ZZ
WWLPHR	*	87.95	-5.46	-3.12	88.39	-5.12	-3.08	ZZ
YNKCC9		94.08	0.67	0.38	94.01	0.50	0.30	ZZ
Z72MUZ		94.16	0.75	0.43	94.21	0.70	0.42	ZZ
Z9KY9W		94.13	0.72	0.41	93.68	0.17	0.10	ZZ

<b>Summary Statistics</b>	<u>Sample GN97</u>	<u>Sample GN98</u>
<b>Grand Means</b>	93.41 Percent	93.51 Percent
<b>Std Dev Btwn Labs</b>	1.75 Percent	1.66 Percent
Statistics based on 26 of 26 reporting participants.		

**Key to Instrument Codes Reported by Participants**

ZZ Instruments No Longer Tracked



# Paper & Paperboard Interlaboratory Testing Program

Report #3152G,  
December 2021


## Analysis 384

Opacity (89% Reflectance Backing) - Fine Papers

TAPPI Official Test Method T425

Grand Mean Sample GN97 = 93.407  
Percent

Grand Mean Sample GN98 = 93.506  
Percent





**Paper & Paperboard Interlaboratory Testing Program**  
**Analysis 386**  
**Opacity (Paper Backing) - Fine Papers and Newsprint**  
**TAPPI Official Test Method T519**

**Report #3152G,**  
**December 2021**

WebCode	Data Flag	<u>Sample GP97</u>			<u>Sample GP98</u>			Instr Code
		Lab Mean	Diff from Grand Mean	CPV	Lab Mean	Diff from Grand Mean	CPV	
4FGQCB		94.99	0.62	1.89	94.52	0.22	1.04	ZZ
677PQA		94.39	0.02	0.07	94.44	0.13	0.63	ZZ
C27VQK		94.18	-0.19	-0.58	94.37	0.07	0.33	ZZ
CHD83B		94.18	-0.19	-0.59	94.31	0.00	0.00	ZZ
CKEBRC		94.60	0.23	0.69	94.51	0.20	0.97	ZZ
CZV789		94.47	0.10	0.32	94.26	-0.04	-0.20	ZZ
LNPB4V		93.90	-0.47	-1.43	93.90	-0.41	-1.94	ZZ
W83JPD		94.25	-0.12	-0.37	94.13	-0.18	-0.83	ZZ

<b>Summary Statistics</b>	<u>Sample GP97</u>	<u>Sample GP98</u>
<b>Grand Means</b>	94.37 Percent	94.31 Percent
<b>Std Dev Btwn Labs</b>	0.33 Percent	0.21 Percent
Statistics based on 8 of 8 reporting participants.		

**Key to Instrument Codes Reported by Participants**

ZZ Instruments No Longer Tracked



# Paper & Paperboard Interlaboratory Testing Program

Report #3152G,  
December 2021


## Analysis 386

Opacity (Paper Backing) - Fine Papers and Newsprint

TAPPI Official Test Method T519

Grand Mean Sample GP97 = 94.370  
Percent

Grand Mean Sample GP98 = 94.305  
Percent



If fewer than 20 laboratories are included in an analysis, a control ellipse will not be drawn on the two-sample plot.



**Paper & Paperboard Interlaboratory Testing Program**  
**Analysis 390**  
**Directional Brightness**  
**TAPPI Official Test Method T452**

Report #3152G,  
December 2021

WebCode	Data Flag	Sample GR97			Sample GR98			Instr Code
		Lab Mean	Diff from Grand Mean	CPV	Lab Mean	Diff from Grand Mean	CPV	
3MHNL9		85.77	0.42	0.20	86.19	0.75	0.35	XC
64EQCB		87.45	2.10	1.00	87.46	2.02	0.95	TP
66CZ8E		86.08	0.73	0.35	85.89	0.45	0.21	TT
6BEGCT		82.32	-3.03	-1.45	82.27	-3.17	-1.50	PP
8ZTP8F		85.24	-0.11	-0.05	85.06	-0.38	-0.18	TS
93BB92		85.66	0.31	0.15	85.55	0.11	0.05	HZ
9Z4W3V		84.93	-0.42	-0.20	84.93	-0.52	-0.24	TS
AYP2JZ		85.33	-0.02	-0.01	85.39	-0.05	-0.02	TP
B74AK3	*	87.60	2.25	1.07	88.42	2.98	1.41	HG
HR2UTG		85.70	0.35	0.17	85.75	0.31	0.15	TP
KGXD4N		85.39	0.04	0.02	85.51	0.07	0.03	TT
KP2XVH	*	78.16	-7.19	-3.44	78.35	-7.09	-3.35	TD
NRYBDR		85.66	0.31	0.15	85.61	0.17	0.08	TS
PXKMQZ	X	68.19	-17.16	-8.21	68.41	-17.03	-8.04	TS
PYQCDF		85.14	-0.21	-0.10	85.15	-0.29	-0.14	XX
RW3KQ6		85.34	-0.01	-0.01	85.48	0.04	0.02	TT
RWMWXQ		85.22	-0.13	-0.06	85.26	-0.18	-0.09	HG
TNK9C4		86.81	1.46	0.70	86.84	1.39	0.66	HG
WWLPHR		88.35	3.00	1.43	88.35	2.91	1.37	PE
YNKCC9		84.99	-0.36	-0.17	85.45	0.01	0.00	TS
YQC9YG		85.87	0.52	0.25	85.91	0.46	0.22	HG

Summary Statistics	Sample GR97	Sample GR98
<b>Grand Means</b>	85.35 Percent	85.44 Percent
<b>Std Dev Btwn Labs</b>	2.09 Percent	2.12 Percent
Statistics based on 20 of 21 reporting participants.		

**Comments on Assigned Data Flags for Test #390**

PXKMQZ (X) - Extreme Data.

**Key to Instrument Codes Reported by Participants**

HG	Hunter Labscan / XE	HZ	Hunter Lab ColorFlex EZ Series
PE	Photovolt 577	PP	Technidyne Profile/Plus
TD	Technidyne Color Touch 45X	TP	Technidyne Test/Plus
TS	Technidyne Brighttimer Micro S-5	TT	Technidyne Brighttimer Micro S4-M
XC	X-Rite Color i5	XX	Instrument make/model not specified by lab




**Paper & Paperboard Interlaboratory Testing Program**  
**Analysis 390**  
**Directional Brightness**  
**TAPPI Official Test Method T452**

**Report #3152G,**  
**December 2021**

**Grand Mean Sample GR97 = 85.349**  
**Percent**

**Grand Mean Sample GR98 = 85.441**  
**Percent**

**ANALYSIS 390**





**Paper & Paperboard Interlaboratory Testing Program**  
**Analysis 391**  
**Directional Brightness of Fluorescent Samples**  
**TAPPI Official Test Method T452**

Report #3152G,  
December 2021

WebCode	Data Flag	<u>Sample GZ97</u>			<u>Sample GZ98</u>			Instr Code
		Lab Mean	Diff from Grand Mean	CPV	Lab Mean	Diff from Grand Mean	CPV	
2F6DUG		95.26	-3.23	-2.54	94.97	-3.42	-2.61	TS
8ZTP8F		99.08	0.59	0.47	98.90	0.51	0.39	TS
AYP2JZ		99.03	0.54	0.42	98.81	0.42	0.32	PP
BHWQA4		99.18	0.69	0.54	99.20	0.81	0.62	TS
D9MQ6U		99.42	0.93	0.73	99.72	1.33	1.01	TT
ERTN2V		97.17	-1.33	-1.04	97.09	-1.30	-0.99	LE
H7ZZKQ		99.06	0.57	0.45	98.86	0.47	0.36	PP
HCNDUR		99.03	0.54	0.42	98.79	0.40	0.30	TS
LKRR3X		99.48	0.99	0.78	99.54	1.15	0.88	TT
LZHWAQ		97.32	-1.17	-0.92	97.58	-0.81	-0.62	PP
PQLXBG		99.30	0.81	0.64	98.92	0.53	0.41	TS
Z72MUZ		98.57	0.08	0.06	98.30	-0.09	-0.07	TS

<b>Summary Statistics</b>	<u>Sample GZ97</u>	<u>Sample GZ98</u>
<b>Grand Means</b>	98.49 Percent	98.39 Percent
<b>Std Dev Btwn Labs</b>	1.27 Percent	1.31 Percent

Statistics based on 12 of 12 reporting participants.

**Key to Instrument Codes Reported by Participants**

LE	L & W Elrepho	PP	Technidyne Profile/Plus
TS	Technidyne Brightimeter Micro S-5	TT	Technidyne Brightimeter Micro S4-M




**Paper & Paperboard Interlaboratory Testing Program**  
**Analysis 391**  
**Directional Brightness of Fluorescent Samples**  
**TAPPI Official Test Method T452**

**Report #3152G,**  
**December 2021**

**Grand Mean Sample GZ97 = 98.493**  
**Percent**

**Grand Mean Sample GZ98 = 98.390**  
**Percent**



If fewer than 20 laboratories are included in an analysis, a control ellipse will not be drawn on the two-sample plot.



**Paper & Paperboard Interlaboratory Testing Program**  
**Analysis 392**  
**Diffuse Brightness**  
**TAPPI Official Test Method T525**

Report #3152G,  
December 2021

WebCode	Data Flag	Sample GR97			Sample GR98			Instr Code
		Lab Mean	Diff from Grand Mean	CPV	Lab Mean	Diff from Grand Mean	CPV	
64EQCB		85.69	0.04	0.13	85.69	0.07	0.22	TC
66CZ8E		85.75	0.09	0.31	85.72	0.10	0.32	EG
677PQA		85.81	0.15	0.51	85.78	0.15	0.51	TC
6QD44U		85.78	0.12	0.40	85.85	0.23	0.76	LE
B2TW34		85.59	-0.06	-0.21	85.52	-0.10	-0.33	TC
CHD83B		85.72	0.07	0.23	85.63	0.01	0.02	LA
CKEBRC		85.21	-0.44	-1.49	85.23	-0.39	-1.31	AC
CZV789		85.53	-0.13	-0.42	85.48	-0.15	-0.49	LE
HR2UTG		85.76	0.11	0.37	85.73	0.10	0.35	LT
JKCVED	*	84.81	-0.85	-2.84	84.80	-0.82	-2.76	LE
LNPB4V		85.62	-0.04	-0.12	85.65	0.02	0.08	LE
NKF6GL	X	68.53	-17.12	-57.42	68.80	-16.82	-56.26	TC
NRYBDR		85.67	0.02	0.06	85.65	0.02	0.08	TC
NUQ4DH		85.64	-0.01	-0.05	85.63	0.00	0.01	TC
TNK9C4	*	85.94	0.29	0.96	85.62	-0.01	-0.02	TC
TQTEWK		85.99	0.33	1.11	85.95	0.32	1.08	TC
WUXKG8		85.51	-0.15	-0.49	85.48	-0.14	-0.47	XX
YNKCC9		86.11	0.46	1.53	86.21	0.58	1.95	LT

Summary Statistics	Sample GR97	Sample GR98
<b>Grand Means</b>	85.65 Percent	85.62 Percent
<b>Std Dev Btwn Labs</b>	0.30 Percent	0.30 Percent

Statistics based on 17 of 18 reporting participants.

**Comments on Assigned Data Flags for Test #392**

NKF6GL (X) - Extreme Data.

**Key to Instrument Codes Reported by Participants**

AC	ACS Spectro-Sensor II	EG	Datacolor Elrepho 450X
LA	L & W Elrepho - Autoline	LE	L & W Elrepho
LT	L & W Elrepho SE 071	TC	Technidyne Color Touch Series
XX	Instrument make/model not specified by lab		




**Paper & Paperboard Interlaboratory Testing Program**  
**Analysis 392**  
**Diffuse Brightness**  
**TAPPI Official Test Method T525**

**Report #3152G,**  
**December 2021**

**Grand Mean Sample GR97 = 85.654**  
**Percent**

**Grand Mean Sample GR98 = 85.623**  
**Percent**



If fewer than 20 laboratories are included in an analysis, a control ellipse will not be drawn on the two-sample plot.



**Paper & Paperboard Interlaboratory Testing Program**  
**Analysis 394**  
**Fluorescent Component of Directional Brightness**  
**TAPPI Official Test Method T452**

**Report #3152G,**  
**December 2021**

WebCode	Data Flag	<u>Sample GZ97</u>			<u>Sample GZ98</u>			Instr Code
		Lab Mean	Diff from Grand Mean	CPV	Lab Mean	Diff from Grand Mean	CPV	
2F6DUG		8.328	0.384	0.90	8.278	0.351	0.87	TS
8ZTP8F		8.254	0.310	0.72	8.250	0.323	0.80	TS
AYP2JZ		8.186	0.242	0.57	8.172	0.245	0.60	XX
D9MQ6U		7.520	-0.424	-0.99	7.520	-0.407	-1.00	TT
ERTN2V		8.216	0.272	0.64	8.210	0.283	0.70	LE
H7ZZKQ		7.840	-0.104	-0.24	7.880	-0.047	-0.11	PP
HCNDUR		7.952	0.008	0.02	7.938	0.011	0.03	TS
LKRR3X		7.500	-0.444	-1.04	7.400	-0.527	-1.30	TT
LZHWAQ		7.400	-0.544	-1.27	7.450	-0.477	-1.17	PP
PQLXBG		8.696	0.752	1.76	8.586	0.659	1.62	TS
Z72MUZ		7.494	-0.450	-1.05	7.508	-0.419	-1.03	TS

<b>Summary Statistics</b>	<u>Sample GZ97</u>	<u>Sample GZ98</u>
<b>Grand Means</b>	7.94 Percent	7.93 Percent
<b>Std Dev Btwn Labs</b>	0.43 Percent	0.41 Percent
Statistics based on 11 of 11 reporting participants.		

**Key to Instrument Codes Reported by Participants**

LE	L & W Elrepho	PP	Technidyne Profile/Plus
TS	Technidyne Brightimeter Micro S-5	TT	Technidyne Brightimeter Micro S4-M
XX	Instrument make/model not specified by lab		




**Paper & Paperboard Interlaboratory Testing Program**  
**Analysis 394**  
**Fluorescent Component of Directional Brightness**  
**TAPPI Official Test Method T452**

**Report #3152G,**  
**December 2021**

**Grand Mean Sample GZ97 = 7.9442**  
**Percent**

**Grand Mean Sample GZ98 = 7.9265**  
**Percent**



If fewer than 20 laboratories are included in an analysis, a control ellipse will not be drawn on the two-sample plot.



**Paper & Paperboard Interlaboratory Testing Program**  
**Analysis 395**  
**Specular Gloss at 75 Degrees - High Range**  
**TAPPI Official Test Method T480**

**Report #3152G,**  
**December 2021**

WebCode	Data Flag	<u>Sample GT97</u>			<u>Sample GT98</u>			Instr Code
		Lab Mean	Diff from Grand Mean	CPV	Lab Mean	Diff from Grand Mean	CPV	
64EQCB		71.59	-1.40	-0.88	71.80	-1.17	-0.60	GM
66CZ8E		74.47	1.48	0.93	75.13	2.16	1.11	TH
7C7ME6		73.98	0.99	0.62	72.78	-0.19	-0.10	VM
9Z4W3V		71.81	-1.18	-0.74	71.50	-1.47	-0.76	LA
CKEBRC		75.87	2.88	1.81	76.75	3.78	1.94	LB
DMNHU6		72.68	-0.31	-0.19	73.18	0.21	0.11	LG
E38ZUJ		76.00	3.01	1.89	75.96	2.99	1.54	LF
HR2UTG		70.87	-2.12	-1.33	69.82	-3.15	-1.62	GA
LZHWAQ		72.78	-0.21	-0.13	73.20	0.23	0.12	PP
P3DM6V		71.65	-1.34	-0.84	72.31	-0.66	-0.34	XX
RW3KQ6		73.41	0.42	0.26	73.36	0.39	0.20	TH
RWMWXQ		71.83	-1.16	-0.72	71.76	-1.21	-0.62	PP
YQC9YG		71.85	-1.14	-0.71	70.65	-2.32	-1.20	PP
Z72MUZ		73.03	0.04	0.03	73.43	0.46	0.23	LF

<b>Summary Statistics</b>	<u>Sample GT97</u>	<u>Sample GT98</u>
<b>Grand Means</b>	72.99 Gloss Units	72.97 Gloss Units
<b>Std Dev Btwn Labs</b>	1.60 Gloss Units	1.94 Gloss Units
Statistics based on 14 of 14 reporting participants.		

**Key to Instrument Codes Reported by Participants**

<b>GA</b> BYK-Gardner (model not specified)	<b>GM</b> BYK-Gardner micro-gloss
<b>LA</b> L & W Gloss - Autoline 300	<b>LB</b> L & W Gloss Tester Code 224
<b>LF</b> L & W Autoline 400	<b>LG</b> L & W Autoline 600
<b>PP</b> Technidyne Profile/Plus	<b>TH</b> Technidyne T480A
<b>VM</b> Valmet PaperLab (was Kajaani/Robotest)	<b>XX</b> Instrument make/model not specified by lab




**Paper & Paperboard Interlaboratory Testing Program**  
**Analysis 395**  
**Specular Gloss at 75 Degrees - High Range**  
**TAPPI Official Test Method T480**

**Report #3152G,**  
**December 2021**

**Grand Mean Sample GT97 = 72.987**  
**Gloss Units**

**Grand Mean Sample GT98 = 72.974**  
**Gloss Units**

**ANALYSIS 395**



If fewer than 20 laboratories are included in an analysis, a control ellipse will not be drawn on the two-sample plot.



**Paper & Paperboard Interlaboratory Testing Program**  
**Analysis 396**  
**Specular Gloss at 75 Degrees - Low Range**  
**TAPPI Official Test Method T480**

**Report #3152G,**  
**December 2021**

WebCode	Data Flag	<u>Sample GU97</u>			<u>Sample GU98</u>			Instr Code
		Lab Mean	Diff from Grand Mean	CPV	Lab Mean	Diff from Grand Mean	CPV	
3MHNL9		46.20	-2.48	-0.83	47.29	-2.88	-0.88	TH
4JW9MZ		46.54	-2.14	-0.71	53.07	2.90	0.88	GM
93BB92		52.78	4.10	1.37	52.64	2.47	0.75	GS
9DCURG		43.92	-4.76	-1.59	49.00	-1.17	-0.36	WJ
CKEBRC		48.50	-0.18	-0.06	50.15	-0.02	-0.01	LA
JKCVED		50.32	1.64	0.55	43.79	-6.38	-1.95	TH
KGXD4N		52.14	3.45	1.15	47.25	-2.92	-0.89	TH
NKF6GL		51.67	2.99	1.00	52.21	2.04	0.62	TH
QX2Y7K		46.05	-2.63	-0.88	52.28	2.11	0.64	PP
TNK9C4		48.69	0.01	0.00	54.01	3.84	1.17	PP

<b>Summary Statistics</b>	<u>Sample GU97</u>	<u>Sample GU98</u>
<b>Grand Means</b>	48.68 Gloss Units	50.17 Gloss Units
<b>Std Dev Btwn Labs</b>	2.99 Gloss Units	3.28 Gloss Units
Statistics based on 10 of 10 reporting participants.		

**Key to Instrument Codes Reported by Participants**

<b>GM</b>	BYK-Gardner micro-gloss	<b>GS</b>	BYK-Gardner Glossgard II
<b>LA</b>	L & W Gloss - Autoline 300	<b>PP</b>	Technidyne Profile/Plus
<b>TH</b>	Technidyne T480A	<b>WJ</b>	Zehntner ZLR 1020




**Paper & Paperboard Interlaboratory Testing Program**  
**Analysis 396**  
**Specular Gloss at 75 Degrees - Low Range**  
**TAPPI Official Test Method T480**

**Report #3152G,**  
**December 2021**

**Grand Mean Sample GU97 = 48.681**  
**Gloss Units**

**Grand Mean Sample GU98 = 50.169**  
**Gloss Units**



If fewer than 20 laboratories are included in an analysis, a control ellipse will not be drawn on the two-sample plot.



**Paper & Paperboard Interlaboratory Testing Program**  
**Analysis 398**  
**Grammage (Mass per Unit Area)**  
**TAPPI Official Test Method T410**

Report #3152G,  
December 2021

WebCode	Data Flag	Sample GW97			Sample GW98			Instr Code
		Lab Mean	Diff from Grand Mean	CPV	Lab Mean	Diff from Grand Mean	CPV	
2F6DUG		104.6	1.2	1.89	75.77	0.54	1.13	ZZ
2W76TW	*	103.6	0.1	0.24	76.38	1.15	2.39	ZZ
2YXZF7		103.5	0.0	0.03	75.64	0.41	0.86	ZZ
3MHNL9		103.4	0.0	-0.03	75.62	0.39	0.81	ZZ
4FGQCB		102.9	-0.6	-0.92	74.75	-0.48	-1.00	ZZ
6JUQ92		104.1	0.6	1.00	75.86	0.63	1.31	ZZ
9DCURG		103.5	0.0	0.02	75.14	-0.10	-0.20	ZZ
AA6XFR	*	103.9	0.4	0.69	74.19	-1.04	-2.16	ZZ
B74AK3		103.2	-0.2	-0.37	75.33	0.10	0.21	ZZ
BHWQA4		103.5	0.1	0.09	75.04	-0.19	-0.39	ZZ
CKEBRC		103.2	-0.2	-0.38	75.00	-0.23	-0.48	ZZ
CZV789		104.5	1.0	1.64	75.50	0.27	0.56	ZZ
G74YQU		103.5	0.0	0.05	75.19	-0.04	-0.09	ZZ
GHXCN7		103.7	0.3	0.49	75.29	0.06	0.12	ZZ
GXATV8		102.8	-0.7	-1.08	74.54	-0.69	-1.43	ZZ
HGKCXP		103.0	-0.4	-0.68	75.19	-0.04	-0.09	ZZ
HWT43E		104.6	1.1	1.85	75.72	0.49	1.02	ZZ
JKCVED		103.0	-0.5	-0.77	75.20	-0.03	-0.06	ZZ
K3UAC3		104.1	0.6	1.00	75.72	0.49	1.02	ZZ
KGXD4N		104.1	0.7	1.07	75.58	0.35	0.72	ZZ
NKF6GL		102.6	-0.9	-1.41	74.36	-0.87	-1.81	ZZ
NUQ4DH	X	5.2	-98.3	-160.03	3.78	-71.45	-148.35	ZZ
PJMK6L		102.4	-1.0	-1.70	74.66	-0.57	-1.18	ZZ
PYCQDF	X	112.1	8.6	14.02	82.35	7.12	14.78	ZZ
QHMZT7		103.4	-0.1	-0.14	75.08	-0.15	-0.31	ZZ
TD4HUL		103.5	0.0	0.05	75.32	0.09	0.18	ZZ
TYLRVJ		103.7	0.2	0.40	75.31	0.08	0.17	ZZ
VUHHUG		103.5	0.0	0.06	74.80	-0.43	-0.89	ZZ
W83JPD		102.7	-0.8	-1.26	75.04	-0.19	-0.39	ZZ
Z9KY9W		102.3	-1.1	-1.83	75.22	-0.01	-0.02	ZZ

Summary Statistics	Sample GW97	Sample GW98
<b>Grand Means</b>	103.44 g/sq m	75.23 g/sq m
<b>Std Dev Btw Labs</b>	0.61 g/sq m	0.48 g/sq m
Statistics based on 28 of 30 reporting participants.		

**Comments on Assigned Data Flags for Test #398**

- NUQ4DH (X) - Extreme Data.
- PYCQDF (X) - Extreme Data.



**Paper & Paperboard Interlaboratory Testing Program**  
**Analysis 398**  
**Grammage (Mass per Unit Area)**  
**TAPPI Official Test Method T410**

**Report #3152G,**  
**December 2021**

**Key to Instrument Codes Reported by Participants**

**ZZ** Instruments No Longer Tracked



# Paper & Paperboard Interlaboratory Testing Program

Report #3152G,  
December 2021


## Analysis 398

Grammage (Mass per Unit Area)

TAPPI Official Test Method T410

Grand Mean Sample GW97 = 103.44  
g/sq m

Grand Mean Sample GW98 =  
75.230 g/sq m





**Paper & Paperboard Interlaboratory Testing Program**  
**Analysis 399**  
**Sizing Test (Hercules Type)**  
**TAPPI Official Test Method T530**

Report #3152G,  
December 2021

WebCode	Data Flag	Sample GX97			Sample GX98			Instr Code
		Lab Mean	Diff from Grand Mean	CPV	Lab Mean	Diff from Grand Mean	CPV	
4JW9MZ		13.79	2.80	1.11	13.16	2.20	0.96	HE
4W48F8		10.29	-0.70	-0.28	11.08	0.12	0.05	HE
677PQA		8.46	-2.53	-1.01	8.05	-2.91	-1.26	HE
6BEGCT	X	16.67	5.68	2.26	12.49	1.53	0.67	HE
7C7ME6		9.68	-1.31	-0.52	9.72	-1.24	-0.54	HE
7WAFCW		9.80	-1.19	-0.47	10.20	-0.76	-0.33	HE
8ZTP8F		10.72	-0.27	-0.11	10.46	-0.50	-0.22	HE
AJVLBU		10.48	-0.51	-0.20	10.73	-0.23	-0.10	HE
AYP2JZ		16.62	5.63	2.24	15.82	4.86	2.12	HE
BHWQA4		8.90	-2.09	-0.83	10.10	-0.86	-0.37	HE
CM9YCT		9.77	-1.22	-0.49	8.96	-2.00	-0.87	HE
D9MQ6U		9.00	-1.99	-0.79	9.40	-1.56	-0.68	HE
GUCGTF		11.36	0.37	0.15	10.94	-0.02	-0.01	HE
H7ZZKQ	X	20.93	9.94	3.95	21.76	10.80	4.70	HE
HCNDUR		10.36	-0.63	-0.25	11.13	0.17	0.08	HE
HTEGBV		7.43	-3.56	-1.42	7.25	-3.71	-1.61	HE
NCLYLY		10.81	-0.18	-0.07	10.39	-0.57	-0.25	HE
NKF6GL		14.27	3.28	1.30	14.78	3.82	1.66	HE
NRYBDR		13.21	2.22	0.88	12.55	1.59	0.69	HE
P3DM6V		11.05	0.06	0.02	9.95	-1.01	-0.44	HE
PQLXBG		11.75	0.76	0.30	11.93	0.97	0.42	HE
PYCQDF		11.67	0.68	0.27	12.63	1.67	0.73	XX
QX2Y7K		13.57	2.58	1.02	13.40	2.44	1.06	HE
TQTEWK		8.82	-2.17	-0.86	9.08	-1.88	-0.82	HE
TYLRVJ		10.22	-0.77	-0.31	9.72	-1.24	-0.54	HE
WWLPHR	*	17.88	6.89	2.74	16.55	5.59	2.44	HE
WYUZCA		8.18	-2.81	-1.12	9.11	-1.85	-0.80	HE
YNKCC9		8.31	-2.68	-1.07	7.97	-2.99	-1.30	HE
Z72MUZ		10.40	-0.59	-0.24	10.74	-0.22	-0.09	HE

Summary Statistics	Sample GX97	Sample GX98
<b>Grand Means</b>	10.99 Seconds	10.96 Seconds
<b>Std Dev Btwn Labs</b>	2.52 Seconds	2.30 Seconds
Statistics based on 27 of 29 reporting participants.		

**Comments on Assigned Data Flags for Test #399**

6BEGCT (X) - Inconsistent in testing between samples. Inconsistent within the determinations of sample GX97.

H7ZZKQ (X) - Data for both samples are high. Possible Systematic Error. Inconsistent within the determinations of both samples.



**Paper & Paperboard Interlaboratory Testing Program**

**Report #3152G,  
December 2021**

**Analysis 399**

**Sizing Test (Hercules Type)**

**TAPPI Official Test Method T530**

**Key to Instrument Codes Reported by Participants**

HE Hercules Sizing Tester

XX Instrument make/model not specified by lab




**Paper & Paperboard Interlaboratory Testing Program**  
**Analysis 399**  
**Sizing Test (Hercules Type)**  
**TAPPI Official Test Method T530**

**Report #3152G,**  
**December 2021**

**Grand Mean Sample GX97 = 10.993**  
**Seconds**

**Grand Mean Sample GX98 = 10.956**  
**Seconds**

**ANALYSIS 399**



-End of Report-